



SUPPLEMENTAL/BID BULLETIN
ADDENDUM NO. 19-007 (Goods)

This Addendum No. **19-007 (Goods)** is issued to modify or amend items in the Bid Documents – with Solicitation No. 2019-01-009 which was posted in the PHILGEPSS with Reference No. 5938179. This shall form an integral part of the Bid Documents.

Name of Project : Procurement for the Supply and Delivery of LABORATORY SUPPLIES AND EQUIPMENT
 EMB-NCB-GOODS-09-2019

INVITATION TO BID (page 4):

Old	New
5. A complete set of Bidding Documents may be acquired by interested Bidders on December 28, 2018 to January 18, 2019, 8:00 a.m. to 2:00 p.m. from the address below and upon payment of a nonrefundable fee for the Bidding Documents, pursuant to the latest Guidelines issued by the GPPB, in the amount of Php5,000.00 .	5. A complete set of Bidding Documents may be acquired by interested Bidders on December 28, 2018 to January 18, 2019, 8:00 a.m. to 2:00 p.m. from the address below and upon payment of a nonrefundable fee for the Bidding Documents, pursuant to the latest Guidelines issued by the GPPB, in the amount of Php1,300.00 .

TECHNICAL SPECIFICATIONS (page 69):

Old	New
EXPLOSION/FIRE PROOF CABINET -For internal use. -body made of sheet steel plastic laminated -Adjustable shelving. -Built-in Spill Tray. -Lockable Doors with cylinder lock and with locking state indicator -colour: body anthracite gray -Self closing doors with hold open facility (doors close automatically in fire). exhaust air connection on top of the cabinet -2 level shelves -fire resistance 90 minutes	EXPLOSION/FIRE PROOF CABINET -For internal use. -body made of sheet steel plastic laminated -Adjustable shelving. -Built-in Spill Tray. -Lockable Doors with cylinder lock and with locking state indicator -colour: body anthracite gray/ yellow -Self closing doors with hold open facility (doors close automatically in fire). exhaust air connection on top of the cabinet -2 level shelves Additional specifications: -freestanding cabinet (915x460x1830mm)
Note: suppliers must be exclusive distributor	Note: suppliers must have a certificate of distributorship with aftersales

For guidance and information of all concerned.


ENGR. MA. CORAZON G. DACUYAN
BAC Chairperson

Received by the Bidder:

Date: _____