



SUPPLEMENTAL BID BULLETIN
ADDENDUM NO. 19-010 (Goods)

This Addendum No. **19-010 (Goods)** is issued to modify or amend items in the Bidding Documents – with Solicitation No. 2019-03-085 which was posted in the PHILGEPS with Reference No. 6109500. This shall form an integral part of the Bid Documents.

Name of Project : Procurement for the Supply and Delivery of One (1) Unit Acid Steam Cleaning System
EMB-NCB-GOODS-14-2019


Section VI. Schedule of Requirements (page 70):

Old	New
Delivery period within forty five (45) calendar days upon receipt of Notice to Proceed.	Delivery period within sixty (60) calendar days upon receipt of Notice to Proceed.

Section VII. Technical Specifications (page 73):

Old	New
SUPPLY AND DELIVERY OF ONE (1) UNIT ACID STEAM CLEANING SYSTEM Technical Requirements: <ul style="list-style-type: none">- stainless steel compartment- max temperature: 220°C- with exhaust/cooling device- with flask for drainage- glass vessels for acids- with initial supply of acids- with tube connections- direct temperature control- used for glasswares- automatic filling/draining system Includes aftersales, 1 year warranty on parts & services Note: Supplier must have certificate of distributorship	SUPPLY AND DELIVERY OF ONE (1) UNIT ACID STEAM CLEANING SYSTEM Technical Requirements: <ul style="list-style-type: none">- stainless steel compartment- max temperature: 220°C- with exhaust/cooling device- with flask for drainage- glass vessels for acids- with tube connections- direct temperature control- used for glasswares- automatic filling/draining system Includes aftersales, 1 year warranty on parts & services Note: Supplier must have certificate of distributorship

For guidance and information of all concerned.


ENGR. REYNALDO B. BARRA
BAC Chairperson

Received by the Bidder:

Date: _____